

UKCA – Declaration of Conformity

We declare that the product Temperature Calibrator

series

TP 17450

manufactured by SIKA Dr. Siebert & Kühn GmbH & Co. KG

comply with

Electrical Equipment (Safety) Regulations 2016

Electromagnetic Compatibility Regulations 2016

The Restriction of the Use of Certain Hazardous Substances in

Electrical and Electronic Equipment Regulations 2012

The devices comply with following technical specifications

BS EN 61010-1:2010+

A1:2019

Safety requirements for electrical equipment for measurement,

control, and laboratory use. General requirements

BS EN IEC 61010-2-

010:2020

Safety requirements for electrical equipment for measurement, control and laboratory use. Particular requirements for laboratory

equipment for the heating of materials

BS EN 61326-1:2013-02 Electrical equipment for measurement, control and laboratory use.

EMC requirements. General requirements

BS EN IEC 63000:2018 Technical documentation for the assessment of electrical and

electronic products with respect to the restriction of hazardous

substances

Kaufungen, 08.07.2021

le lelette

Dipl.- Ing. K. Ulloth

(Manager CE- Coordination and Safety Supervisor)

(CE- Koordinator und Produktsicherheitsbeauftragter)