



UKCA – Declaration of Conformity

We declare that the product **Temperature Indicator**
series **DiTemp LCK**
manufactured by **SIKA Dr. Siebert & Kühn GmbH & Co. KG**
comply with

Electromagnetic Compatibility Regulations 2016

The Restriction of the Use of Certain Hazardous Substances in
Electrical and Electronic Equipment Regulations 2012

The devices comply with following technical specifications

- | | |
|----------------------|--|
| BS EN 61326-1:2013 | Electrical equipment for measurement, control and laboratory use.
EMC requirements. General requirements |
| BS EN IEC 63000:2018 | Technical documentation for the assessment of electrical and
electronic products with respect to the restriction of hazardous
substances |

Kaufungen, 26.01.2023

i.V. 

Dipl.- Ing. K. Ulloth
(Manager CE- Coordination and Safety Supervisor)
(CE- Koordinator und Produktsicherheitsbeauftragter)